

File :Q:\GCMS-08 Minimal\DATA\20221103\48 -jjp.D
Operator : Organic QC Lab
Acquired : 04 Nov 2022 01:18 using AcqMethod Hy_Pest_15cm_S20.M
Instrument : GCMS 8
Sample Name: AE-00017
Misc Info : 100-500 ug/ml
Vial Number: 7



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

