

File : B:\QC Data\GCMS-04 Minimal\DATA\092719\20-.D  
Operator : KM/ZA  
Acquired : 28 Sep 2019 00:06 using AcqMethod VOC-20SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name : AE-00034-CAL-01  
Misc Info :  
Vial Number: 10



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

