

File : B:\QC Data\GCMS-04 Minimal\DATA\092719\26-.D
Operator : KM/ZA
Acquired : 28 Sep 2019 2:47 using AcqMethod VOC-20SPLIT DECKER.M
Instrument : GCMS4
Sample Name : AE-00034-CAL-04
Misc Info :
Vial Number : 13



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

