

File :Q:\GCMS-04 Minimal\DATA\102219\35.D
Operator : RM / KM / ZA
Acquired : 23 Oct 2019 2:29 using AcqMethod VOC-50SPLIT DECKER.M
Instrument : GCMS4
Sample Name: AE-0004-CAL-03
Misc Info :
Vial Number: 21



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

