

File :L:\QC Data\GCMS-07 Minimal\DATA\06062201\35 - SAK.D
Operator : Organic QC department
Acquired : 06 Jun 2022 22:30 using AcqMethod KERLA-2Split-2uL.M
Instrument : 5977B MSD
Sample Name:AE-00041-10ML
Misc Info :
Vial Number: 40



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

