


File :D:\MassHunter\GCMS\1\DATA\12142101\17-KD.D
Operator : Organic QC department
Acquired : 14 Dec 21 02:20 pm using AcqMethod KERLA-VOC-SV_S100.M
Instrument : 5977B MSD
Sample Name: ALR-007S-ET-10X
Misc Info :
Vial Number: 42

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

