

File :D:\MassHunter\GCMS\1\DATA\050721\02- ALR.D  
Operator : Organic QC Lab  
Acquired : 07 May 2021 10:04 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name: 221051141 (ALR-099S)  
Misc Info : Dicyclohexyl Ohthalate in Methanol  
Vial Number: 1



**AccuStandard®**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

