

=====

Acq. Operator	: SYSTEM	Seq. Line	: 5
Acq. Instrument	: LCMS-1	Location	: P2-B5
Injection Date	: 4/2/2021 7:15:14 AM	Inj	: 1
		Inj Volume	: 1.000 µl
Acq. Method	: C:\CHEM32\1\METHODS\Purity195_315.M		
Last changed	: 5/8/2020 11:20:44 AM by SYSTEM		
Analysis Method	: C:\CHEM32\1\METHODS\Purity195_315.M		
Last changed	: 3/31/2021 3:04:04 PM by SYSTEM		
	(modified after loading)		
Method Info	: Operator: Oraganic QC lab		
	Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.		

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

