



```

=====
Acq. Operator   : SYSTEM                      Seq. Line : 12
Acq. Instrument : LCMS-1                     Location  : P1-B2
Injection Date  : 11/2/2023 2:15:34 PM      Inj       : 1
                                              Inj Volume: 1.000 µl
  
```

```

Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 1/5/2023 12:01:53 PM by SYSTEM
Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.
  
```

Sample Info : Monobenzyl phthalate in ACN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

