

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    1
Acq. Instrument : LCMS-1                      Location  : P1-C8
Injection Date  : 11/4/2021 3:05:09 PM       Inj       :    1
                                                Inj Volume: 1.000 µl

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/29/2021 2:16:13 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/8/2021 11:40:47 AM by SYSTEM
Method Info     : Operator: Organic QC lab
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.
  
```



Sample Info : Monoethylhexyl phthalate @ 100 ug/mL in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

