

File :D:\MassHunter\GCMS\1\DATA\010422\19- ALR.D
Operator : Organic QC Lab
Acquired : 04 Jan 2022 15:43 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: ALR-172S
Misc Info : b-Ionone in MeOH
Vial Number: 32



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

