

File :Q:\GCMS-07 Minimal\DATA\07102001\15- YX.D
Operator : Organic QC department
Acquired : 10 Jul 2020 21:54 using AcqMethod KERLA-2017_S100.M
Instrument : 5977B MSD
Sample Name: APP-9-013
Misc Info : Anthracene in MeOH
Vial Number: 131



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

