

File :Q:\GCMS-04 Minimal\DATA\040622A\39- YX.D
Operator : Organic QC Lab
Acquired : 7 Apr 2022 8:37 using AcqMethod VOC-100SPLIT# .M
Instrument : GCMS4
Sample Name: APP-9-039
Misc Info : Chlorobenzene in MeOH
Vial Number: 28



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

