

File : L:\QC Data\GCMS-04 Minimal\DATA\122019\46-.D  
Operator :  
Acquired : 21 Dec 2019 6:00 using AcqMethod VOC-100SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name : APP-9-048-R1-50X  
Misc Info : CHLOROPRENE  
Vial Number: 23



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

