

File : L:\QC Data\GCMS-09 Minimal\DATA\11012022\33 =SAK

... =.D

Operator :

Instrument : GCMS 9

Acquired : 01 Nov 2022 20:55 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split Time 12 min .M

Sample Name : APP-9-050

Misc Info :



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

