

File :D:\MassHunter\GCMS\1\data\09121901\19- .D
Operator : KD
Acquired : 12 Sep 2019 16:48 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: APP-9-076
Misc Info :
Vial Number: 24



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

