

File :D:\MassHunter\GCMS\1\DATA\01172001\10- D
Operator : Organic QC department
Acquired : 17 Jan 20 11:17 am using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: APP-9-086
Misc Info : a,a-Dimethylphenethylamine in DCM
Vial Number: 128



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

