

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-09 Minimal\DATA\09132023  
... \27- KM.D  
Operator :  
Instrument : GCMS 9  
Acquired : 13 Sep 2023 20:48 using AcqMethod ALR 19 min 20 split FAST .M  
Sample Name: APP-9-089  
Misc Info : m-Dinitrobenzene in DCM



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

