

File :D:\MassHunter\GCMS\1\DATA\030122\11- ALR.D
Operator : Organic QC Lab
Acquired : 01 Mar 2022 11:27 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: APP-9-148-20X
Misc Info : N-Nitrosodiethylamine in DCM
Vial Number: 20



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

