

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    3
Acq. Instrument : LCMS-1                      Location  : P1-C6
Injection Date  : 3/19/2020 11:09:51 AM      Inj       :    1
                                                    Inj Volume: 1.000 µl
  
```

```

Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/18/2019 4:18:14 PM by SYSTEM
Method Info     : Operator: GA
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                  :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                  particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                  MS Detector. No UV Reference.
  
```

```

Additional Info : Peak(s) manually integrated
  
```

## Current Chromatogram(s)

