



File :L:\QC Data\GCMS-06 Minimal\DATA\042723\22 -SAK.D  
Operator : Organic QC Lab  
Acquired : 27 Apr 2023 17:53 using AcqMethod CHICK\_2019\_VOC-SV.M  
Instrument : GCMS 6  
Sample Name: APP-9-214  
Misc Info :  
Vial Number: 41

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

