

File :D:\MassHunter\GCMS\1\data\20230131\08- -JP-.D
Operator : Organic QC Lab
Acquired : 31 Jan 23 12:22 pm using AcqMethod Hy_VOC-SV_24cm_S100.M
Instrument : GCMS 8
Sample Name: AS-E0020
Misc Info : p-Chloro-m-cresol; in meoh
Vial Number: 2



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

