

File :Q:\GCMS-06 Minimal\DATA\071720\23- ZA.D  
Operator : Organic QC Lab  
Acquired : 17 Jul 2020 15:41 using AcqMethod M502-FAST.M  
Instrument : GCMS 6  
Sample Name: AS-E0055  
Misc Info : o-Nitrophenol in MeOH  
Vial Number: 111



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

