

File :Q:\GCMS-08 Minimal\DATA\20210413\21- YX.D
Operator : Organic QC Lab
Acquired : 13 Apr 2021 17:41 using AcqMethod Hy_Pest_15cm_S100.M
Instrument : GCMS 8
Sample Name: AS-E0067
Misc Info : Di-n-octyl phthalate in MeOH
Vial Number: 115



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25 μ m
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1 μ l inj.; GC/MS; Injector 270°C

