

File :Q:\GCMS-08 Minimal\DATA\20191021\34- .D
Operator : KM / JP / ZA
Acquired : 22 Oct 2019 03:34 using AcqMethod KERLA-2017_S100.M
Instrument : GCMS 8
Sample Name: AS-E0068
Misc Info : Diethyl phthalate ug/mL in MeOH
Vial Number: 40



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

