

File : Z:\QC Data\GCMS-07 Minimal\DATA\12181901\22-.D  
Operator : Organic QC department  
Acquired : 18 Dec 2019 16:54 using AcqMethod KERLA-2017.M  
Instrument : 5977B MSD  
Sample Name: AS-E0071  
Misc Info : in Acetone  
Vial Number: 137



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

