

File :Q:\GCMS-08 Minimal\DATA\20220208\22- =JP=.D
Operator : Organic QC Lab
Acquired : 08 Feb 22 05:54 pm using AcqMethod Hy_PAH_24cm_S100.M
Instrument : GCMS 8
Sample Name: AS-E0078
Misc Info : Fluorene; in meoh
Vial Number: 74



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

