

File :Q:\GCMS-04 Minimal\DATA\051519\BSB\09-.D  
Operator : KM  
Acquired : 15 May 2019 19:42 using AcqMethod VOC-100SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name: AS-E0083  
Misc Info :  
Vial Number: 27



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

