

File :Q:\GCMS-08 Minimal\DATA\20220606\20- [JP].D
Operator : Organic QC Lab
Acquired : 06 Jun 2022 15:29 using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: AS-E0089
Misc Info : Chlordane (Tech Mix); in meoh
Vial Number: 56



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

