

File : Q:\GCMS-08 Minimal\DATA\20220729\2 -jopo.D
Operator : Organic QC Lab
Acquired : 29 Jul 2022 15:31 using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: AS-E0129-3
Misc Info : Aroclor 1260; in iso
Vial Number: 5



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

