

File :L:\QC Data\GCMS-08 Minimal\DATA\20220523\16 SAK.D
Operator : Organic QC Lab
Acquired : 23 May 2022 13:02 using AcqMethod Hy_PAH_24cm_S100.M
Instrument : GCMS 8
Sample Name: AS-E0129-4
Misc Info :
Vial Number: 29

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

