

File :K:\QC Data\GCMS-04 Minimal\DATA\062420\49RM.D  
Operator : ORGANIC QC DEPT.  
Acquired : 25 Jun 2020 12:29 using AcqMethod VOC-100SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name: AS-E0201  
Misc Info :  
Vial Number: 42



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

