

File : M:\GCMS-06 Minimal\DATA\112221\07- ALR.D  
Operator : Organic QC Lab  
Acquired : 22 Nov 2021 17:28 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name : AS-E0231  
Misc Info : Dibenz(a,h)anthracene in p-Dioxane  
Vial Number: 68



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

