



AccuStandard®

File :M:\GCMS-07 Minimal\DATA\05272101\27- ALR.D  
Operator : Organic QC department  
Acquired : 27 May 2021 16:21 using AcqMethod KERLA-2017\_Short\_15Min\_S100.M  
Instrument : 5977B MSD  
Sample Name: AS-E0298  
Misc Info : N-Nitrosopyrrolidine in Methanol  
Vial Number: 33

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

