

=====
Acq. Operator : SYSTEM Seq. Line : 7
Acq. Instrument : LCMS-1 Location : P1-D2
Injection Date : 5/30/2019 11:56:55 AM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : D:\DATA\05301901\05301901 2019-05-30 10-04-33\Purity195_315.M

Last changed : 5/30/2019 10:04:33 AM by SYSTEM

Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M

Last changed : 4/10/2019 10:56:50 AM by SYSTEM

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

