

File :Q:\GCMS-08 Minimal\DATA\20201019\17- (JP).D
Operator : Organic QC Lab
Acquired : 19 Oct 2020 16:28 using AcqMethod Hy_BDE_S1.M
Instrument : GCMS 8
Sample Name: BDE-116S
Misc Info : 2,3,4,5,6-Penta-BDE in isooctane
Vial Number: 59



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

