

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20230404
... \BSB\54- KM.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 05 Apr 2023 00:29 using AcqMethod Hy_BDE209_S1.M
Sample Name: BDE-205S
Misc Info : 2,3,3',4,4',5,5',6-Octabromodiphenyl ether in Isooctane



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

