

File :Q:\GCMS-08 Minimal\DATA\20190506\14-.D  
Operator : JP/KM  
Acquired : 06 May 2019 22:15 using AcqMethod KERLA-2017\_S200.M  
Instrument : GCMS 8  
Sample Name: BF-SOY-ME  
Misc Info : AOCS Standard  
Vial Number: 99



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

