

File : L:\QC Data\GCMS-06 Minimal\DATA\012023\20 - sak.D
Operator : Organic QC Lab
Acquired : 20 Jan 2023 17:09 using AcqMethod CHICK_2019_13MIN_2split_2uL.M
Instrument : GCMS 6
Sample Name : C-004S-TP
Misc Info : diluted to approx. 35 ug/mL
Vial Number : 129



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

