

File :D:\MassHunter\GCMS\1\data\12072022\22- ALR.D  
Operator :  
Acquired : 07 Dec 2022 14:20 using AcqMethod Hy SV DB-5MS UI20M 0.18 2 split FAST .M  
Instrument : GCMS 9  
Sample Name: C-010S  
Misc Info : RT: 2,6-Dichlorobiphenyl in Isooctane  
Vial Number: 35



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

