

File : L:\QC Data\GCMS-06 Minimal\DATA\040623\02 -SAK.D
Operator : Organic QC Lab
Acquired : 06 Apr 2023 16:56 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name : C-019S-TP
Misc Info :
Vial Number : 35



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

