

File :Q:\GCMS-08 Minimal\DATA\20210811\35- _JP_.D
Operator : Organic QC Lab
Acquired : 12 Aug 21 08:55 pm using AcqMethod Hy_PCB_10cm_S1.M
Instrument : GCMS 8
Sample Name: C-046S-TP
Misc Info : 22'36'-TetraPCB; in iso
Vial Number: 57



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

