

File : Q:\GCMS-07 Minimal\DATA\05282001\22- -DC.D
Operator : Organic QC department
Acquired : 28 May 20 03:34 pm using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name : C-047S-TP
Misc Info : 2,2',4,4'-Tetrachlorobiphenyl in Isooctane
Vial Number: 72



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

