

File :Q:\GCMS-06 Minimal\DATA\120121\16- KM.D
Operator : Organic QC Lab
Acquired : 02 Dec 2021 09:52 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: C-048S
Misc Info : 2,2',4,5-Tetrachlorobiphenyl in Isooctane
Vial Number: 99



AccuStandard[®]

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25 μ m
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1 μ l inj.; GC/MS; Injector 270°C

