

File :Q:\GCMS-06 Minimal\DATA\011320\11-.D
Operator : Organic QC Lab
Acquired : 13 Jan 2020 17:07 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: C-054S-TP
Misc Info : 2,2',6,6'-Tetrachlorobiphenyl in Isooctane
Vial Number: 30



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

