

File :Q:\GCMS-08 Minimal\DATA\20210525\53- [ JP ].D  
Operator : Organic QC Lab  
Acquired : 26 May 2021 00:44 using AcqMethod Hy\_Pest\_15cm\_S10.M  
Instrument : GCMS 8  
Sample Name: C-070S-TP  
Misc Info : 2,3',4',5-TetraPCB; in isooctane  
Vial Number: 84



Column: DB-5, 30m, 0.25ID, 0.25um  
Oven: 80c, 2m, 30c/m to 320, final 2min  
GC para: Pulse split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj, GCMS, Inj 280c

