

File :Q:\GCMS-07 Minimal\DATA\05192001\33- JP.D  
Operator : Organic QC department  
Acquired : 20 May 2020 09:21 using AcqMethod KERLA-2017.M  
Instrument : 5977B MSD  
Sample Name: C-098S-TP  
Misc Info : 2,2',3',4,6-Penta-PCB in isoctane  
Vial Number: 95



Column: DB-5, 30m, 0.25ID, 0.25um  
Oven: 80c, 2m, 30c/m to 320, final 2min  
GC para: Pulse split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj, GCMS, Inj 280c

