

File : Q:\GCMS-08 Minimal\DATA\20210820\62- }JP{.D  
Operator : Organic QC Lab  
Acquired : 21 Aug 2021 22:28 using AcqMethod Hy\_PCB\_5cm\_S5.M  
Instrument : GCMS 8  
Sample Name: C-1115-TP  
Misc Info : 233'55'-Penta-PCB in isoctane  
Vial Number: 86



Column: DB-5, 30m, 0.25ID, 0.25um  
Oven: 80c, 2m, 30c/m to 320, final 2min  
GC para: Pulse split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj, GCMS, Inj 280c

