

File :Q:\GCMS-08 Minimal\DATA\20220720\27- -JP).D
Operator : Organic QC Lab
Acquired : 20 Jul 22 06:25 pm using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: C-126S (222071331)
Misc Info : 33'44'5-PentaPCB; in iso
Vial Number: 34



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

