


File :D:\MassHunter\GCMS\1\data\20230906\06 - (JP).D  
Operator : Organic QC Lab  
Acquired : 06 Sep 23 02:03 pm using AcqMethod Hy\_Pest\_24cm\_S2\_2ulinj.M  
Instrument : GCMS 8  
Sample Name: C-136S  
Misc Info : 22'33'66'-Hexa-PCB in iso  
Vial Number: 4

 **AccuStandard®**  
Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

